


N25 DIGITAL PANEL METERS

FEATURES:



- Destined for measurement of d.c. voltage or d.c. current, temperature through Pt100 resistance thermometers, J, K thermocouples, a.c. voltage and a.c. current.
- 5 LED digit displays with 14 mm digit high.
- Parameters programmable by PD14 programmer:
 - precision of displayed results (decimal point),
 - measurement averaging time,
 - recounting of indications (individual characteristic),
 - automatic or manual compensation: cold junction temperature for measurement with thermocouples or wire resistance for measurement with Pt100 (N25T).


INPUTS:




OUTPUTS:



GALVANIC ISOLATION:



EXAMPLE OF APPLICATION



Measurement and display:
 - temperature
 - analog signals
 - d.c. current and voltage
 - rms current and voltage.

INPUTS

Type	Measuring ranges		Parameters	Overloads	Errors				
N25S	$-11 \text{ mV} \dots -10 \text{ mV} \dots 60 \text{ mV} \dots 66 \text{ mV}$		Input resistance $> 1 \text{ M}\Omega$	Short duration overload (1s): - voltage input: 10 Un - current input: 5 In Sustained overload: 110% Un, 110% In	Basic error: $\pm (0.2\% \text{ of range} + 1 \text{ digit})$ Additional error from ambient temperature changes: $\pm (50\% \text{ of basic error}/10K)$				
	$-66 \text{ mV} \dots 60 \text{ mV} \dots 60 \text{ mV} \dots 66 \text{ mV}$								
	$-0.5 \text{ V} \dots 0 \text{ V} \dots 10 \text{ V} \dots 11 \text{ V}$								
	$-11 \text{ V} \dots -10 \text{ V} \dots 10 \text{ V} \dots 11 \text{ V}$		Input resistance $10 \Omega \pm 1\%$						
	$-1 \text{ mA} \dots 0 \text{ mA} \dots 20 \text{ mA} \dots 22 \text{ mA}$								
N25T	$3.6 \text{ mA} \dots 4 \text{ mA} \dots 20 \text{ mA} \dots 22 \text{ mA}$		Input resistance $10 \Omega \pm 1\%$	Short duration overload (1s) Input of sensors: 30 V	Basic error: $\pm (0.2\% \text{ of range} + 1 \text{ digit})$ Additional errors: <ul style="list-style-type: none"> compensation of cold junction temperature changes: $\pm 0.2\% \text{ of range}$, from ambient temperature changes: $\pm (50\% \text{ of basic error}/10K)$. 				
	Pt100	$-50^\circ\text{C} \dots 150^\circ\text{C}$	Current flowing through the sensor: $< 300 \mu\text{A}$. Resistance of wires connecting RTD with the meter: - max 5Ω (per wire) for automatic compensation - max 10Ω (per wire) for manual compensation						
		$-50^\circ\text{C} \dots 400^\circ\text{C}$							
	Thermo-couple J	$-50^\circ\text{C} \dots 1200^\circ\text{C}$							
N25Z	Thermo-couple K	$-50^\circ\text{C} \dots 1370^\circ\text{C}$		Short term overload (1s): voltage input: 2 Un ($< 1000\text{V}$), current input: 10 In Sustained overload: 150% Un (for 400V input), 120% (for remaining inputs), 120% In	Basic error: <ul style="list-style-type: none"> voltage and current: $\pm (0.5\% \text{ of range} + 1 \text{ digit})$ in frequency range $20 \dots 500 \text{ Hz}$ frequency: $\pm (0.02\% \text{ of range} + 1 \text{ digit})$ Additional error from ambient temperature changes: $\pm (50\% \text{ of basic error}/10K)$				
	$1 \dots 100 \dots 120 \text{ V a.c.}$								
	$2.5 \dots 250 \dots 300 \text{ V a.c.}$								
	$4 \dots 400 \dots 600 \text{ V a.c.}$								
	$20 \dots 500 \text{ Hz}$ (in voltage range: $24 \dots 480 \text{ V}$)								
	$0.01 \dots 1 \dots 2 \text{ A a.c.}$		Input resistance $10 \text{ m}\Omega \pm 10\%$						
N25H	$0.05 \dots 5 \dots 6 \text{ A a.c.}$		Input resistance $2 \text{ m}\Omega \pm 10\%$						
	$0 \dots 100 \dots 110 \text{ V d.c.}$		Input resistance $> 2 \text{ M}\Omega$	Short term overload (1s): voltage input: 2 Un ($< 1000\text{V}$), current input: 10 In Sustained overload: 150% Un (for $\pm 400\text{V}$ input), 120% (for remaining inputs), 120% In	Basic error: $\pm (0.2\% \text{ of range} + 1 \text{ digit})$ Additional error from ambient temperature changes: $\pm (50\% \text{ of basic error}/10K)$				
	$0 \dots 250 \dots 275 \text{ V d.c.}$								
	$-120 \dots -100 \dots 100 \dots 120 \text{ V d.c.}$								
	$-300 \dots -250 \dots 250 \dots 300 \text{ V d.c.}$								
	$-600 \dots -400 \dots 400 \dots 600 \text{ V d.c.}$								
	$-1.2 \dots -1 \dots 1 \dots 2 \text{ A d.c.}$		Input resistance $10 \text{ m}\Omega \pm 10\%$						
	$-6 \dots -5 \dots 5 \dots 6 \text{ A d.c.}$		Input resistance $2 \text{ m}\Omega \pm 10\%$						

OUTPUTS

For N25S and N25T	Output for supply external transducers	$24 \text{ V} \pm 5\%, 30 \text{ mA}$
-------------------	--	---------------------------------------

Sifam Tinsley Instrumentation Ltd
 Unit 1 Warner Drive,
 Springwood Industrial Estate
 Braintree, Essex, UK, CM7 2YW
 E-mail: sales@sifamtinsley.com
 Web: www.sifamtinsley.com/uk
 Contact: +44(0)1803 615139



Sifam Tinsley Instrumentation Inc.
 3105 Creekside Village Drive,
 Suite No. 801, Kennesaw,
 Georgia 30144 (USA)
 E-mail Id : psk@sifamtinsley.com
 Web: www.sifamtinsley.com
 Contact No.: +1 404 736 4903

EXTERNAL FEATURES

Weight	< 0.25 kg	
Overall dimensions	96 x 48 x 64 mm with terminals	
Protection grade (acc. to EN 60529)	ensured by the housing: IP65	from the terminal side: IP 20
Display	5-digit LED display, 14 mm high, red colour	indication range: -19999...99999

RATED OPERATING CONDITIONS

Supply voltage	230 V ± 10% a.c. (45...65 Hz); 110 V ± 10% a.c. (45...65 Hz) 24 V ± 10% a.c. (45...65 Hz); 85...253 V a.c. (40...400 Hz) or d.c.; 20...40 V a.c. (40...400 Hz) or d.c.	input power consumption: 6 VA
Temperature	ambient: -10...23...55°C	storage: -25...85 °C
Relative humidity	≤ 95%	condensation inadmissible
Operating position	any	
Preheating time	30 min	
Averaging time	≥ 0.5 s	1 second default set

SAFETY AND COMPATIBILITY REQUIREMENTS

Electromagnetic compatibility	noise immunity noise emissions	acc. to EN 61000-6-2 acc. to EN 61000-6-4
Isolation between circuits	basic	
Pollution grade	2	
Installation category	III (for the 400 V option - category II)	
Maximal phase-to-earth operating voltage	for supply circuits: 300 V, for measuring circuits: 600 V - cat. II for other circuits: 50 V	acc. to EN 61010-1
Altitude above sea level	< 2000 m	

CONNECTION DIAGRAMS




Fig. 1. Electrical connections of the N25S meter




Fig. 2. Electrical connections of the N25T meter




Fig. 3. Connections of N25T measuring inputs




Fig. 4. Electrical connections of N25Z and N25H meters for the measurement of voltage (and frequency only in N25Z)




Fig. 5. Electrical connections of N25Z i N25H meters for the current measurement

ORDERING

TABLE 1. ORDERING CODES:

N25 -	X	X	X	XX	XX	X	X
Input kind:							
standard: voltage, current	S						
temperature: thermocouples, resistance thermometers	T						
a.c. signals	Z						
d.c. signals: high voltage and high current	H						
Input: see table 2			X				
Supply:							
230 V a.c.			1				
110 V a.c.			2				
24 V a.c.			3				
85...253 V a.c./d.c. with supply output 24 V/30 mA*			4				
20...40 V a.c./d.c. with supply output 24 V/30 mA*			5				
Unit: see table 3			XX				
Version:				00			
standard							
non-standard settings				NS			
custom-made**				XX			
Language:					P		
Polish					E		
English					X		
other**							
Acceptance tests:							
without extra requirements			0				
with an extra quality inspection certificate			1				
acc. to customer's request**			X				

* - the output is only in N25S and N25T meters

** - after agreeing with the manufacturer

TABLE 2. INPUT SIGNALS

Nr	N25S	N25T	N25Z	N25H
1	0...20 mA	Pt100: -50...150°C	100 V a.c.	±100 V d.c.
2	4...20 mA	Pt100: -50...400°C	250 V a.c.	±250 V d.c.
3	0...60 mV	Thermocouple J	400 V a.c.	±400 V d.c.
4	0...10 V	Thermocouple K	1 A a.c.	±1 A d.c.
5	± 60 mV		5 A a.c.	±5 A d.c.
6	± 10 V		20...500 Hz	0...100 V d.c.
7				0...250 V d.c.

TABLE 3. CODES OF PRINTED UNITS:

Code	Unit	Code	Unit	Code	Unit
00	without unit	06	mA	12	bar
01	°C	07	kA	13	kPa
02	%	08	kV	14	MPa
03	A	09	Hz		
04	V	10	turns	XX	on order
05	mV	11	rpm		

TABLE 4. EXAMPLE OF NON-STANDARD SETTINGS:

Parameter	Range/Value
Decimal point	000,0 for I, U
Averaging time	1 s
Upper measurement overflow	99999
Lower measurement overflow	-19999
Individual characteristic	enabled
Parameter a of the individual characteristic	5
Parameter b of the individual characteristic	0
Order example 1 : The code N25Z-2 1 04 0 E 0 means: N25Z - digital meter for d.c. signals	
2 - input: 250 V a.c.	
1 - supply: 230 V a.c.	
04 - unit: V	
0 - standard version	
E - English language	
0 - without extra requirements	
Order example 2 : The code N25S-1 4 02 E 1 means: N25S - digital meter for d.c. signal	
1 - input: 0...20mA	
4 - supply: 85...253 V a.c.	
02 - unit: %	
NS - non-standard settings, display range: 0...100.0	
E - English language	
1 - with an extra quality inspection certificate	

Sifam Tinsley Instrumentation Ltd
Unit 1 Warner Drive,
Springwood Industrial Estate
Braintree, Essex, UK, CM7 2YW
E-mail: sales@sifamtinsley.com
Web: www.sifamtinsley.com/uk
Contact: +44(0)1803 615139



Sifam Tinsley Instrumentation Inc.
3105, Creekside Village Drive,
Suite No. 801, Kennesaw,
Georgia 30144 (USA)
E-mail Id : psk@sifamtinsley.com
Web: www.sifamtinsley.com
Contact No.: +1 404 736 4903